1-1-1982

Measurement of Synchrotron x-ray energies and line shapes using diffraction markers

Juana Vivó Acrivos

San Jose State University, juana.acrivos@sjsu.edu

S.S. P. Parkin

San Jose State University

K. Hathaway

San Jose State University

J. R. Reynolds

San Jose State University

M. P. Klein

University of California, Berkeley

See next page for additional authors

Follow this and additional works at: https://scholarworks.sjsu.edu/chem_pub

Part of the [Physical Chemistry Commons](#)

Recommended Citation

Juana Vivó Acrivos, S.S. P. Parkin, K. Hathaway, J. R. Reynolds, M. P. Klein, A. Thompson, and D. Goodin. "Measurement of Synchrotron x-ray energies and line shapes using diffraction markers" *Review of Scientific Instruments* (1982): 575-581. <https://doi.org/10.1063/1.1137026>

This Article is brought to you for free and open access by the Chemistry at SJSU ScholarWorks. It has been accepted for inclusion in Faculty Publications, Chemistry by an authorized administrator of SJSU ScholarWorks. For more information, please contact scholarworks@sjsu.edu.

Authors

Juana Vivó Acrivos, S.S. P. Parkin, K. Hathaway, J. R. Reynolds, M. P. Klein, A. Thompson, and D. Goodin

Measurement of synchrotron xray energies and line shapes using diffraction markers

J. V. Acrivos, K. Hathaway, J. Reynolds, J. Code, S. Parkin et al.

Citation: [Rev. Sci. Instrum.](#) **53**, 575 (1982); doi: 10.1063/1.1137026

View online: <http://dx.doi.org/10.1063/1.1137026>

View Table of Contents: <http://rsi.aip.org/resource/1/RSINAK/v53/i5>

Published by the [American Institute of Physics](#).

Related Articles

ECE-imaging of the H-mode pedestal (invited)

[Rev. Sci. Instrum.](#) **83**, 10E329 (2012)

Response to "Comment on 'Scalings for radiation from plasma bubbles'" [
[Phys. Plasmas](#) **18**, 034701 (2011)

]
[Phys. Plasmas](#) **18**, 034702 (2011)

Betatron x-ray generation from electrons accelerated in a plasma cavity in the presence of laser fields

[Phys. Plasmas](#) **16**, 103103 (2009)

Dynamics of emitting electrons in strong laser fields

[Phys. Plasmas](#) **16**, 093115 (2009)

Undulator radiation in a periodic magnetic field with a constant component

[J. Appl. Phys.](#) **104**, 124507 (2008)

Additional information on Rev. Sci. Instrum.

Journal Homepage: <http://rsi.aip.org>

Journal Information: http://rsi.aip.org/about/about_the_journal

Top downloads: http://rsi.aip.org/features/most_downloaded

Information for Authors: <http://rsi.aip.org/authors>

ADVERTISEMENT

AIPAdvances

Special Topic Section:
PHYSICS OF CANCER

Why cancer? Why physics? [View Articles Now](#)

Measurement of synchrotron x-ray energies and line shapes using diffraction markers

J. V. Acrivos, K. Hathaway, J. Reynolds, J. Code, and S. Parkin

San José State University, San José, California 95192

M. P. Klein, A. Thompson, and D. Goodin

Laboratory of Chemical Biodynamics, Lawrence Berkeley Laboratory, University of California, Berkeley, California 94720

(Received 20 August 1981; accepted for publication 18 January 1982)

Standard reference markers for accurate, reproducible synchrotron x-ray energies are obtained using a three Si crystal spectrometer. The first two crystals are in the monochromator and the third is used to obtain diffraction markers which monitor the energy. Then for any value of the glancing angle on the reference Si crystal the energy for the (333) diffraction must occur at $3/4$ that of the (444) and $3/5$ of that for the (555). This establishes for the first time an absolute synchrotron energy scale. Higher-order diffractions are used to determine excitation line profiles. We conclude that the use of reference diffractions is necessary to measure reproducible x-ray energies and to analyze the incident photons' line profile. The detection of diffractions near the edge of measurement and near the Cu edge will provide a fast secondary standard which will allow comparison of edge data between different laboratories. The diffraction profiles will allow the proper analysis of spectral line widths.

PACS numbers: 41.70. + t

INTRODUCTION

The accurate measurement of x-ray energies E near 10^4 eV to within 10^2 ppm is needed in order to ascertain the valence of the absorber in conductors and in biological materials because the valence is an indicator of how metal-insulator transitions and how biological red-ox reactions occur. In particular, we have correlated the metal-insulator transitions in the liquid metal Rb-NH₃ with the changes in the x-ray absorption spectra near the Rb K edge.¹ Here a single absorption near the edge ($1s \rightarrow np$, $n \geq 4$) shifts as a function of the relative metal concentration by as much as 5 eV as the valence changes from 0 to +1. Also multiple valence has been deduced from shifts in the x-ray absorption for SmS under pressure² or by Y substitution in Sm_xY_{1-x}S³ and by intercalation in TaSe₂.⁴ The relative changes are easy to measure during a given experiment by running a set of standards together with the unknown samples.⁵ However, in order to compare data taken at different times, at different laboratories, it is important to measure the absolute value of the energy to within at least 10^2 ppm. The purpose of this work is to describe a simple three Si crystal spectrometer shown in Fig. 1 which does the job.

The principles used to measure accurate wavelengths are described in Sec. II.⁶ The experimental details and error analysis for an example are applied in Sec. III to measurements of the TaS₂-N₂H₄ and the C (graphite) + AsF₅ systems carried out at the Stanford Synchrotron Radiation Laboratory (SSRL).

II. X-RAY WAVELENGTHS

Accurate x-ray wavelength measurements are carried out using the Stenström correction to the Bragg relation.⁶ The Bragg relation is satisfied for the planes in the two Si crystals of the monochromator (M) and for an additional reference crystal (g) with respective scattering vectors S when:

$$2 \sin \theta_B / \lambda = S, \quad (1)$$

where θ_B is the angle which gives the correct wavelength λ in air. Then if $\theta_B = \theta_{MB}$ at the monochromator crystals and $\theta_B = \theta_{gB}$ at a third reference crystal in the x-ray beam, a reference diffraction may be recorded when

$$\sin \theta_{MB} / S_M = \sin \theta_{gB} / S_g = \lambda / 2 = hc / 2E. \quad (2)$$

The value of θ_B is determined from the Darwin relation which corrects the glancing angle for the effects of the index of refraction. The x-ray glancing angle and the wavelength change from (θ, λ) in air to (θ', λ') within the crystal and are related by the real part of the index of refraction⁶:

$$1 - \delta = \cos \theta / \cos \theta' = \lambda / \lambda'. \quad (3)$$

The Stenström correction to the energy (due to δ) in relations (2) is then

$$\Delta E_i / E_i = -\delta_i / \sin^2 \theta_i = \Delta_{is}, \quad (4)$$

where $i = g$ or M and near the Cu K edge ($\delta = \delta_0 = 7.3$

FIG. 1. Block diagram of the experimental configuration used to introduce reference markers as SSRL.

$\times 10^{-6}$ and $\theta_B \approx 21^\circ$) this is of the order of 60 ppm or if θ is the measured value⁶:

$$\theta_B = \theta - 2\delta \csc 2\theta. \quad (5)$$

Thus to evaluate the energy to within an accuracy of 10^2 ppm, one needs to know the glancing angle θ to within the same accuracy. If θ can be measured directly the solution is obvious.⁶ The situation with synchrotron sources (e.g., at SSRL), however, is less direct. There the monochromator is driven by a stepping motor (M step) with a known incremental angle per step (r_{θ_M}), but the absolute angle is unknown and must be determined by a calibration (e.g., at a known energy). We may write:

$$\theta = \theta_M + c_M \quad \text{where} \quad \theta_M = r_{\theta_M}(M \text{ steps}) \quad (5')$$

and c_M is an offset angle to be determined by a calibration. The correct energy selected by the monochromator crystal is then given by

$$E_M = hcS_M[1 + \delta/\sin^2(\theta_M + c_M)]/2 \sin(\theta_M + c_M). \quad (6)$$

In a conventional two Si crystal spectrometer, c_M is calibrated by identifying the reported edge energy E_0 with some reasonable feature of the metal x-ray absorption near the Cu K edge. Some users identify the first inflection point with E_0 (8.9803 keV) and one assumes that δ varies linearly with $\lambda^{2.6}$; i.e., the Stenström relation (4) is a constant. However, near the Fe K edge this calibration of c_M places the first inflection point 2 eV above E_0 (7.1112 keV). This is not possible because the true edge is usually preceded in energy by some white line features. Other users place E_0 near the first sharp feature (minimum or maximum) of the extended x-ray absorption fine structure (EXAFS) which is ambiguous but much more plausible than the first inflection point. Also internal standards can be used to provide relative energy measurements.⁵ However, an absolute energy calibration requires a three-crystal spectrometer. Figure 1 describes the first one used successfully at SSRL, but the simple apparatus can be used at any other synchrotron source.

The only variable in the optical system shown in Fig.

1 is the glancing angle θ_0 of x rays (coming through counter I) on a Si crystal (4×10^{-4} m thick) cut along the (111) [or (110)] plane which is changed from 45° by a step motor in order to place two or more diffraction markers near the edge being measured (Fig. 2) and two or more near the Cu K edge. The x-ray diffraction signal is recorded on a third (fluorescence) channel. The energy calibration is carried out by generating tables of E_g vs θ_0 ($\theta_0 = 40^\circ$ – 45° in increments of 0.1°) using relations (2) and (6) and carrying out three additional edge measurements at the beginning and end of every period, during which θ_0 is left unchanged. First, the Cu K edge is run. The strong diffractions near the Cu K edge shown in Fig. 2 are for $g = (333)$ and (335) . These are easily identified and used to determine θ_0 to within 100 ppm by interpolating from the tables. Then the (444) diffraction near the As K - or Ta L_I edge is used to ascertain c_M from the fact that $E_{333}/E_{444} = 3/4$. Finally, the edge being investigated is scanned and the strong diffractions predicted by the tables are assigned. The procedure requires only three extra scans for calibration and the wealth of diffractions for the Si crystal allows one to place a diffraction marker or two near the edge being measured. The diffraction markers are then used to check the stability of the calibration throughout an entire experiment as long as care is exercised to maintain θ_0 constant.

The reference relation (2) is written for the crystallographic planes normal to $g = (hkl)$ in the reference marker using the crystal Cartesian axes ($Z \parallel g_0$, $X \parallel g_x$, and $Y \parallel \epsilon$, the direction of the beam polarization in Fig. 3), i.e.,

$$S_g \sin(\theta_M + c_M) = S_M \sin \theta_0 [g \cdot g_0 / (|g| \cdot |g_0|) + \cot \theta_0 g \cdot g_x / (|g| \cdot |g_x|)] [1 - \Delta_{gs} + \Delta_{Ms}], \quad (2')$$

where both g_0 and g_x are first estimated by a Laue photograph but the exact orientation of the reference crystal in the laboratory axes must be obtained from (2') for a series of calibration points. Also the value of θ_0 for different S_g diffractions has a specific dependence on θ_M ,

FIG. 2. Diffraction markers obtained with a Si crystal with $g_0 = (111)$ and $g_x = (\bar{1}\bar{1}2)$ in Fig. 3; (a) near the Cu-K and (b) Ta- L_I edges.

i.e., from (2') it follows that a given g diffraction tracks θ_M of the monochromator at the rate of

$$\left(\frac{\partial\theta_0}{\partial\theta_M}\right) = \frac{\tan(A_g + \theta_0)}{\tan(\theta_M + c_M)}, \quad (7)$$

where $\tan A_g = |g_0|g \cdot g_x / [|g_x|g \cdot g_0]$ giving a change of energy with $\Delta\theta_0$ of:

$$\Delta E/E = -\cot(A_g + \theta_0)\Delta\theta_0 \quad (8)$$

and from (2') and (8) both θ_0 and A_g can be determined exactly. Figure 4 shows the dependence of θ_0 on θ_M for the diffraction $g = (\bar{4}26)$ in the orientation shown in Fig.

3 with $g_0 = (111)$ and $g_x = (\bar{1}\bar{1}2)$. The errors introduced by alignment are discussed in the Appendix. These do not affect the calibration. The relations (6) used to evaluate c_M are

$$u_{3,4} = 4E_{333}/3E_{444} \\ = 4 \sin(r_{\theta_M}M_4 + c_M)/3 \sin(r_{\theta_M}M_3 + c_M) = 1,$$

and/or

$$u_{3,5} = 5E_{333}/3E_{555} \quad (6') \\ = 5 \sin(r_{\theta_M}M_5 + c_M)/3 \sin(r_{\theta_M}M_3 + c_M) = 1,$$

where the M_g (monochromator) steps are experimental values measured when relation (2) is satisfied for the reference g diffraction. Thus the relations (6') calibrate the synchrotron radiation energy in the manner used earlier by both Stenström and Dumond.⁶

III. EXAMPLE OF WORK DONE AT SSRL

In the system shown in Fig. 1, the SLO-SYN motor rotates at the rate of $0.9^\circ/\text{step}$ advancing the micrometer heat at the rate $r_Z = 1.588 \times 10^{-6}$ m/step causing a change in θ_0 which is determined by the separation between the ball bearings supporting the back and front of the prism crystal mount of 4.58 cm. Then the rate of change is $r_{\theta_0} = -3.47 \times 10^{-5}$ rad/step, and the fractional rate of change in the energy which satisfies the Bragg condition for the g plane follows from Eq. (8), i.e.,

$$r_{Eg} = (\Delta E/E)/\text{step} = -r_{\theta_0} \cot(A_g + \theta_0)/\text{step}. \quad (9)$$

FIG. 3. Diffraction direction for a general crystal plane g .

FIG. 4. Plot of θ_0/step vs θ_M/M step following the $g = (\bar{2}\bar{4}6)$ diffraction (when $g_0 = (111)$ and $g_x = (\bar{1}\bar{1}\bar{2})$ in Fig. 3).

This is to be compared with the fractional rate of energy change per monochromator step:

$$r_{EM} = (\Delta E/E)/M \text{ Step} \\ = -r_{\theta_M} \cot(\theta_M + c_M)/M \text{ Step}, \quad (9')$$

where at the SSRL Station I-5 used in this work $r_{\theta_M} = 1.0908 \times 10^{-6} \text{ rad}/M \text{ step}$. The slope measured in Fig. 4 is in agreement with the calculated value in Table I:

$$r = r_{EM}/r_E \\ = -3.144 \times 10^{-2} \tan(A_g + \theta_0) \cot(\theta_M + c_M) = -0.123,$$

indicating that this diffraction marker can be moved at the rate of 0.28 eV per step of the reference motor (g step) so that it can be placed at any desired energy to within a 25 ppm accuracy. Then if θ_0 is left unchanged during a set of measurements the absolute energy can be determined accurately by measuring say the (333), (444), and (555) diffractions. We have found that θ_0 did remain constant for several days of running time but that any disturbance of the Si crystal prism mount can change θ_0 .

The energy calibration is carried out by minimizing the error in relations (6'):

$$\text{DELU}_{i,j} = 10^6 \Delta u_{i,j} = 10^6 |1 - u_{i,j}|, \quad (10)$$

versus c_m as indicated in Fig. 6 and in Table I. The error propagated into $c_m(\Delta c_m)$ depends on the diffractions i, j ; i.e.,

$$\Delta c_m = \{ \Delta u_{i,j} + |\theta_{M_i} \cot \theta_{M_iB} - \theta_{M_j} \cot \theta_{M_jB}| \Delta r_{\theta_M} / r_{\theta_M} \\ + |\theta_{M_i} \cot \theta_{M_iB}| \Delta M_i / M_i + |\theta_{M_i} \cot \theta_{M_iB}| \Delta M_j / M_j \} \\ \times \{ |\cot \theta_{M_jB} - \cot \theta_{M_iB}| \}^{-1} \quad (10')$$

decreases as the denominator increases, giving a sharper minimum for $\Delta u_{3,5}$ than for $\Delta u_{3,4}$ vs c_M in Fig. 6. Therefore we use $u_{3,5}$ for the calibration. Then under experimental conditions where the error in θ_{M_i} is 10 ppm and the diffraction positions are known only to within the Stenström correction (60 ppm), we obtain an error Δc_M

TABLE I. Typical interpolation table used to determine $C_M = 8.6540 \pm 0.002 \text{ deg}$.

M-steps	θ_{MB}/deg	$E_M \text{ keV}$
<i>E₃₃₃</i> data (three measurements)		
211506	21.8731	8.6670
<i>E₄₄₄</i> data (two measurements)		
121091	16.2222	11.5581
<i>E₅₅₅</i> data (two measurements)		
68200	12.9165	14.4449

Errors in ppm: Del $U_{3,4} = 185$, Del $U_{3,5} = 2$

Calibration data: $E_{001/2}$, E_{335} , 1st inflection points near Fe, Cu K edges, Ta LIII and L_I edges

308577	27.9401	6.8913
203867	21.3957	8.8510
294702	27.0729	7.0945
200000	21.1540	8.9474
168500	19.1853	9.8255
119727	16.1369	11.6175

Typical table of energies evaluated using relation (6) with $C_M = 8.6540 \text{ deg}$
 $\theta_0 = 43.1900 \text{ deg}$, $g_0 = 1.00 \ 1.00 \ 1.00$, $g_x = -1.00 \ -1.00 \ 2.00$

h	k	l	E_g/keV	θ_g/deg	$M\text{-step}/10^3$	θ_{MB}/deg	$dE_g/d\theta_0/\text{eV}/\text{deg}$
0	0	12	2×6.9156	82.0744	306.8681	27.8333	16.8
3	3	3	8.6670	43.1900	211.5032	21.8729	-161.1
3	3	5	8.8651	57.6101	203.2936	21.3598	-98.2
4	4	4	11.5560	43.1900	121.1381	16.2251	-214.9
5	5	5	14.4450	43.1900	68.1990	12.9164	-268.6

FIG. 5. Diffraction line profiles: (a) $g = (\bar{4}\bar{2}\bar{6})$. (b) $g = (333)$ and (335) . The x-ray unit for the g diffraction is $l_g = -[1 + (\Delta E/E)/\Delta_{g,1}]$. For $g = \bar{4}\bar{2}\bar{6}$, the origin of $\Delta E/E$ is given for the Bragg angle θ_B . However, if the line profile is due to the monochromator diffractions then the x-ray unit is l_M , where $(l_M + 1)/(l_g + 1) = 0.1433$ and the linewidth for each diffraction is: $\Delta\theta = \tan(\theta_M + c_M)\Delta E(\text{FWHH})/2E = 3.58 \times 10^{-5} \text{ rad} = 0.74 \text{ s of arc}$.

= 30 ppm (2×10^{-3} DEG). Table I gives the energies evaluated in the neighborhood of the Cu and Fe K edges and the Ta L_I edge to within an accuracy

$$\Delta E/E = \Delta S_M/S_M + [\Delta c_M + \theta_M(\Delta M/M + \Delta r_{\theta_M}/r_{\theta_M})]\cot \theta_{MB}. \quad (10'')$$

However, the advantage of the method described here lies in the constant monitoring of the energy via the diffraction profiles run with every spectrum. The use of reference markers in the study of the $\text{TaS}_2 + \text{N}_2\text{H}_4$ system and $\text{C} + \text{AsF}_5$ is shown in Figs. 5, 7, and 8. The actual width of the diffractions was found to depend on the incident flux. For low incident flux (12 mA current) the full width at half-height (FWHH) of the diffraction in Figs. 5 and 7 is 2.3–2.6 eV, whereas for high incident flux (65 mA current) we observe a FWHH of 4.4 eV which may be due to poor adjustment of slit widths in the monochromator.⁶

The sharp diffractions [$\Delta E(\text{FWHH})/E = 2 \times 10^{-4}$] in Fig. 5 are used as an indication of the reliability of the data. However, this width is considerably greater than the calculated Darwin width for perpendicular polarization⁶:

$$\Delta\theta_g = 4\delta f(g)\csc 2\theta/Z = |\Delta E/E|_{D,g} \tan \theta_B \quad (11)$$

or

$$|\Delta E/E|_{D,g} = 2\delta f(g)\csc^2 \theta_B/Z, \quad (12)$$

where $\delta = 3.8 \times 10^{-6}$ near 11.088 keV for Si, $f(g)$ is the atomic structure factor [$f(220) = 8.70$ and $f(246) = 3.68 \pm 0.037$] and Z is the atomic number. Then the $g = (\bar{2}\bar{4}\bar{6})$, $(\Delta E/E)_{D,g} = 1.1 \times 10^{-5}$ is an order of magnitude smaller than the observed full width at half-height (FWHH) in Fig. 5 of 2×10^{-4} . The Darwin widths are

FIG. 6. Dependence of the minimum in $\Delta u_{i,j}$ vs c_M on the diffractions used for the calibration, $\bullet$: $\Delta u_{3,4}$, $\times$: $\Delta u_{3,5}$.

additive. For the two monochromator crystals ($g = 220$) this is greater than that for the reference; i.e., in relation (12) $2(\Delta E/E)_{D,M} = 1.08 \times 10^{-4}$ near 11.088 keV is of the same order of magnitude as the observed diffraction width. The physical significance here then is that the reference diffraction actually follows the incident photons' line profile.

The error due to the divergence of the synchrotron beam is determined by the slit height $h \approx 10^{-1}$ cm over the distance of the source to slit ($R = 2 \times 10^3$ cm); i.e., $h/R \approx 4 \times 10^{-5}$ rad⁶ is less than the Darwin width in this case. However, the diffractions shown in Fig. 5 do not resemble the Darwin profiles evaluated using relation (6.50) of Ref. 6. They are much more symmetric, suggesting that a detailed investigation of diffraction line profiles using synchrotron radiation can yield information on δ as well as absorption coefficients.

The knowledge of the incident photons' line profile is of some importance in the study of white peaks arising from transitions to bound states below a given edge. In

FIG. 7. Observed energy reproducibility: (a) during the intercalation of TaS₂ to form 3R-TaS₂·I_{4/3}; (b) comparison of edge energies for the three compounds in relation (6).

FIG. 8. Observed energy reproducibility near the As K edge. Here two consecutive runs are superimposed.

the study of IT-TaS₂ we can measure changes in the line profile due to intercalation with N₂H₄. The sharp reference diffractions shown in Fig. 6 (2.6 eV FWHH) ensure that the incident photons' line profiles are narrower than the width of the measured absorption peaks near the Ta-L edges. The shifts in energy shown in Fig. 6 must be related to changes in the structure due to intercalation.^{8,9} Here the intercalation reaction

introduces commensurate periodic lattice distortions into IT-TaS₂ which have been associated with multiple valence states. The accurate measurements shown in Fig. 7 indicate the Ta valence is increased when two layers of N₂H₄ are intercalated in between the TaS₂ layers but that it is decreased when only one layer of N₂H₄ remains in between the layers. These fractional valence changes must be associated with charge transfer from N₂H₄ to the TaS₂ layers.

The width of the white line profile near the As K edge in C(graphite) + AsF₅ shown in Fig. 8 is of the same order of magnitude as the FWHH of the diffraction of 4.4 eV. As a result one cannot determine the actual absorption profile because we suspect the diffraction width is caused by the width of the incident photons' line profile. This poor resolution may be due to heating of the monochromator crystals by the incident flux and/or poor slit adjustment. However, the reference diffraction is the only possible way to establish the stability and condition of the incident photons.

APPENDIX

The $n(111)$ diffractions used to calibrate c_M in the three-crystal spectrometer are insensitive to any misalignment in g_0 , g_x , and g_y . However, since other $g = (hkl)$ diffractions may be used to monitor the energy near an edge being measured, it is important to determine

any alignment errors. Let

$$\begin{aligned} \mathbf{g}_x &= (\bar{1}\bar{1}2) + \mathbf{x}; & \mathbf{g}_y &= (1\bar{1}0) + \mathbf{y}; \\ \mathbf{g}_0 &= (111) + \mathbf{z}, \end{aligned} \quad (\text{A1})$$

where $\mathbf{x}$, $\mathbf{y}$, $\mathbf{z}$ represent alignment errors which are related by the orthogonality relations:

$$\begin{aligned} \mathbf{g}_x \cdot \mathbf{g}_0 &\cong \mathbf{x} \cdot \mathbf{g}_0 + \mathbf{z} \cdot \mathbf{g}_x \cong \mathbf{g}_y \cdot \mathbf{g}_0 \cong \mathbf{y} \cdot \mathbf{g}_0 + \mathbf{z} \cdot \mathbf{g}_y \\ &\cong \mathbf{g}_x \cdot \mathbf{g}_y \cong \mathbf{x} \cdot \mathbf{g}_y + \mathbf{y} \cdot \mathbf{g}_x \cong 0. \end{aligned} \quad (\text{A2})$$

The diffraction energy is evaluated in (2) using the ratio $\sin \theta_g/g$ and the error introduced by a misalignment (A1) into the energy is:

$$\Delta_a g = \Delta E_g/E_g = -(\bar{\theta}_g - \theta_g) \cot \theta_g, \quad (\text{A3})$$

where $\bar{\theta}_g$ is the true value and θ_g assumes $\mathbf{x} = \mathbf{y} = \mathbf{z} = 0$. Then using (2'):

$$\begin{aligned} \Delta_a g &= -(\sin \bar{\theta}_g - \sin \theta_g)/\sin \theta_g \\ &= -\{3^{-3/2} \sin \theta_0 [3\mathbf{z} \cdot \mathbf{g} - (\mathbf{z} \cdot \mathbf{g}_0)(\mathbf{g} \cdot \mathbf{g}_0)] + 6^{-3/2} \\ &\quad \times \cos \theta_0 [6\mathbf{x} \cdot \mathbf{g} - (\mathbf{g}_x \cdot \mathbf{x})(\mathbf{g}_x \cdot \mathbf{g})]\}/(g \sin \theta_g). \end{aligned} \quad (\text{A4})$$

The cubic symmetry of the crystal allows us to simplify the error analysis. Here:

$$\Delta_a nnn = -18^{-1/2} \mathbf{x} \cdot \mathbf{g}_0 \cot \theta_0 \equiv \Delta_a^0,$$

$$\begin{aligned} \Delta_a hhl &= \Delta_a^0 [(2h + l) \\ &\quad - 2^{1/2}(l - h) \tan \theta_0] \sin \theta_0 / (g g_0 \sin \theta_g), \end{aligned} \quad (\text{A4}')$$

e.g., near $\theta_0 = 43^\circ$ we obtain

$$\Delta E_{333}/E_{333} - \Delta E_{335}/E_{335} = 0.41 \Delta_a^0$$

and

$$\Delta E_{333}/E_{333} - \Delta E_{0012}/E_{0012} = 1.1 \Delta_a^0. \quad (\text{A5})$$

When θ_0 is evaluated assuming $\Delta E_{nnn} = 0$, then relations (A5) indicate that the diffractions $\mathbf{g} = (hhl)$ should be predicted to within $\Delta_a^0 = 3.7 \times 10^{-3}$ in Table I.

ACKNOWLEDGMENTS

This work was done under the auspices of NSF DMR 7820577 and proposals 312 M and 512 M at SSRL. The facilities at SSRL are supported under NSF contract DMR 77-27489 in cooperation with the U.S. Department of Energy. We are grateful to J. R. Salem, E. Grove, and T. Rowe for the Si crystals, R. Ott for machine shop work, E. Langone for computer analysis, and S. Wootton, S. Brown, and H. Stein for typing.

- ¹ J. V. Acrivos, K. Hathaway, A. Robertson, A. Thompson, and M. P. Klein, *J. Phys. Chem.* **84**, 1206 (1980); and *J. Phys. C* **14**, 1557 (1981).
- ² J. M. Tranquada, R. Ingalls, J. E. Whitmore, and E. D. Crozier, *Seventh Annual SSRL Meeting*, SSRL Report 80/05, p. 74.
- ³ J. B. Boyce, R. M. Martin and J. W. Allen, *Seventh Annual SSRL Meeting*, SSRL Report 80/05, p. 128.
- ⁴ J. V. Acrivos, S. Parkin, J. Code, J. Reynolds, K. Hathaway, H. Kurasaki, and E. Marseglia, *Seventh Annual SSRL Meeting*, SSRL Report 80/05, p. 79; *J. Phys. C* **14**, L349 (1981).
- ⁵ K. E. Falk, D. B. Goodin, M. P. Klein, and T. Wydrzynski, *Seventh Annual SSRL meeting*, SSRL Report 79/05.
- ⁶ A. H. Compton and S. K. Allison, *X-Rays in Theory and Experiment* (Van Nostrand, Princeton, New Jersey, 1963) Chap. VI; G. S. Brown and S. Doniach, *Synchrotron Radiation*, H. Winnick and S. Doniach (Plenum, New York, 1981).
- ⁷ B. Dawson, *Proc. R. Soc. London* **298A**, 379 (1967).
- ⁸ J. V. Acrivos, *Intercalated Layer Materials*, edited by F. Levy (Reidel, Dordrecht, Holland, 1979), Vol. VI.
- ⁹ G. Tatlock and J. V. Acrivos, *Philos. Mag.* **38**, 81 (1978).